

1 **§291.33 Operational Standards**

2

3

XXX

4 (i) Automated devices and systems.

5 (1) Automated compounding or counting devices. If a pharmacy uses automated compounding
6 or counting devices:

7 (A) the pharmacy shall have a method to calibrate and verify the accuracy of the automated
8 compounding or counting device and document the calibration and verification on a routine
9 basis;

10 (B) the devices may be loaded with bulk or unlabeled drugs only by a pharmacist or by
11 pharmacy technicians or pharmacy technician trainees under the direction and direct
12 supervision of a pharmacist;

13 (C) the label of an automated compounding or counting device container shall indicate the
14 brand name and strength of the drug; or if no brand name, then the generic name, strength, and
15 name of the manufacturer or distributor;

16 (D) records of loading bulk or unlabeled drugs into an automated compounding or counting
17 device shall be maintained to show:

18 (i) name of the drug, strength, and dosage form;

19 (ii) manufacturer or distributor;

20 (iii) manufacturer's lot number;

21 (iv) manufacturer's expiration date;

22 (v) date of loading;

23 (vi) name, initials, or electronic signature of the person loading the automated compounding
24 or counting device; and

25 (vii) signature or electronic signature of the responsible pharmacist; and

26 (E) the automated compounding or counting device shall not be used until a pharmacist
27 verifies that the system is properly loaded and affixes his or her signature to the record as
28 specified in subparagraph (D) of this paragraph.

29 (2) Automated pharmacy dispensing systems.

30 (A) Authority to use automated pharmacy dispensing systems. A pharmacy may use an
31 automated pharmacy dispensing system to fill prescription drug orders provided that:

32 (i) the pharmacist-in-charge is responsible for the supervision of the operation of the system;

33 (ii) the automated pharmacy dispensing system has been tested by the pharmacy and found
34 to dispense accurately. The pharmacy shall make the results of such testing available to the
35 board upon request; and

36 (iii) the pharmacy will make the automated pharmacy dispensing system available for
37 inspection by the board for the purpose of validating the accuracy of the system.

38 (B) Quality assurance program. A pharmacy which uses an automated pharmacy dispensing
39 system to fill prescription drug orders shall operate according to a written program for quality
40 assurance of the automated pharmacy dispensing system which:

41 (i) requires continuous monitoring of the automated pharmacy dispensing system; and

42 (ii) establishes mechanisms and procedures to test the accuracy of the automated
43 pharmacy dispensing system at least every six months and whenever any upgrade or change is
44 made to the system and documents each such activity.

45 (C) Policies and procedures of operation.

46 (i) When an automated pharmacy dispensing system is used to fill prescription drug orders,
47 it shall be operated according to written policies and procedures of operation. The policies and
48 procedures of operation shall:

49 (I) provide for a pharmacist's review, approval, and accountability for the transmission of
50 each original or new prescription drug order to the automated pharmacy dispensing system
51 before the transmission is made;

52 (II) provide for access to the automated pharmacy dispensing system for stocking and
53 retrieval of medications which is limited to licensed healthcare professionals or pharmacy
54 technicians acting under the supervision of a pharmacist;

55 (III) require prior to use, that a pharmacist checks, verifies, and documents that the
56 automated pharmacy dispensing system has been accurately filled each time the system is
57 stocked;

58 (IV) provide for an accountability record to be maintained which documents all transactions
59 relative to stocking and removing medications from the automated pharmacy dispensing
60 system;

61 (V) require a prospective drug regimen review is conducted as specified in subsection
62 (c)(2) of this section; and

63 (VI) establish and make provisions for documentation of a preventative maintenance
64 program for the automated pharmacy dispensing system.

65 (ii) A pharmacy which uses an automated pharmacy dispensing system to fill prescription
66 drug orders shall, at least annually, review its written policies and procedures, revise them if
67 necessary, and document the review.

68 (D) Recovery Plan. A pharmacy which uses an automated pharmacy dispensing system to fill
69 prescription drug orders shall maintain a written plan for recovery from a disaster or any other
70 situation which interrupts the ability of the automated pharmacy dispensing system to provide
71 services necessary for the operation of the pharmacy. The written plan for recovery shall
72 include:

73 (i) planning and preparation for maintaining pharmacy services when an automated
74 pharmacy dispensing system is experiencing downtime;

75 (ii) procedures for response when an automated pharmacy dispensing system is
76 experiencing downtime; and

77 (iii) procedures for the maintenance and testing of the written plan for recovery.

78 (E) Final check of prescriptions dispensed using an automated pharmacy dispensing system.
79 For the purpose of §291.32(c)(2)(D) of this title (relating to Personnel), a pharmacist must
80 perform the final check of all prescriptions prior to delivery to the patient to ensure that the
81 prescription is dispensed accurately as prescribed.

82 (i) This final check shall be considered accomplished if:

83 (I) a check of the final product is conducted by a pharmacist after the automated pharmacy
84 dispensing system has completed the prescription and prior to delivery to the patient; or

85 (II) the following checks are conducted by a pharmacist:

86 (-a-) if the automated pharmacy dispensing system contains bulk stock drugs, a
87 pharmacist verifies that those drugs have been accurately stocked as specified in subparagraph
88 (C)(i)(III) of this paragraph; and

89 (-b-) a pharmacist checks the accuracy of the data entry of each original or new
90 prescription drug order entered into the automated pharmacy dispensing system.

91 (ii) If the final check is accomplished as specified in clause (i)(II) of this subparagraph, the
92 following additional requirements must be met.

93 (I) The dispensing process must be fully automated from the time the pharmacist releases
94 the prescription to the automated pharmacy dispensing system until a completed, labeled
95 prescription ready for delivery to the patient is produced.

96 (II) The pharmacy has conducted initial testing and has a continuous quality assurance
97 program which documents that the automated pharmacy dispensing system dispenses
98 accurately as specified in subparagraphs (A) and (B) of this paragraph.

99 (III) The automated pharmacy dispensing system documents and maintains:

100 (-a-) the name(s), initials, or identification code(s) of each pharmacist responsible for the
101 checks outlined in clause (i)(II) of this subparagraph; and

102 (-b-) the name(s), initials, or identification code(s) and specific activity(ies) of each
103 pharmacist, pharmacy technician, or pharmacy technician trainee who performs any other
104 portion of the dispensing process.

105 (IV) The pharmacy establishes mechanisms and procedures to test the accuracy of the
106 automated pharmacy dispensing system at least every month rather than every six months as
107 specified in subparagraph (B) of this paragraph.

108 (3) Automated checking device.

109 (A) For the purpose of §291.32(c)(2)(D) of this title, the final check of a dispensed prescription
110 shall be considered accomplished using an automated checking device provided:

111 (i) a check of the final product is conducted by a pharmacist prior to delivery to the patient or
112 the following checks are performed by a pharmacist:

113 (I) the prepackaged drug used to fill the order is checked by a pharmacist who verifies that
114 the drug is labeled and packaged accurately; and

115 (II) a pharmacist checks the accuracy of each original or new prescription drug order.

116 (ii) the prescription is dispensed, labeled, and made ready for delivery to the patient in
117 compliance with Class A (Community) Pharmacy rules; and

118 (iii) prior to delivery to the patient:

119 (I) the automated checking device confirms that the correct drug and strength has been
120 labeled with the correct label for the correct patient; and

121 (II) a pharmacist performs all other duties required to ensure that the prescription has been
122 dispensed safely and accurately as prescribed.

123 (B) If the final check is accomplished as specified in subparagraph (A) of this paragraph, the
124 following additional requirements must be met.

125 (i) The pharmacy has conducted initial testing of the automated checking device and has a
126 continuous quality assurance program which documents that the automated checking device
127 accurately confirms that the correct drug and strength has been labeled with the correct label for
128 the correct patient.

129 (ii) The pharmacy documents and maintains:

130 (I) the name(s), initials, or identification code(s) of each pharmacist responsible for the
131 checks outlined in subparagraph (A)(i) of this paragraph; and

132 (II) the name(s) initials, or identification code(s) and specific activity(ies) of each
133 pharmacist or pharmacy technician who perform any other portion of the dispensing process.

134 (iii) The pharmacy establishes mechanisms and procedures to test the accuracy of the
135 automated checking device at least monthly.